



Enfermedades Infecciosas y Microbiología Clínica

www.elsevier.es/eimc



Brief report

Impact of automated nucleic acid extraction platforms on plasma Cytomegalovirus DNA loads quantitated by real-time PCR normalized to the 1st WHO international standard



Ángela Sánchez-Simarro^a, Eliseo Albert^a, Paula Michelena^a, Estela Giménez^{a,b}, David Navarro^{a,b,c,*}

^a Microbiology Service, Clinic University Hospital, INCLIVA Health Research Institute, Valencia, Spain

^b CIBER de Enfermedades Infecciosas, Instituto de Salud Carlos III, Madrid, Spain

^c Department of Microbiology, School of Medicine, University of Valencia, Valencia, Spain

ARTICLE INFO

Article history:

Received 22 May 2024

Accepted 13 July 2024

Available online 3 September 2024

Keywords:

Cytomegalovirus (CMV)

CMV DNA load

Plasma

CMV DNA extraction

Nucleic acid extraction platforms

CMV DNA quantification

ABSTRACT

Introduction: The extent to which commercially available nucleic acid extraction platforms impact the magnitude of Cytomegalovirus (CMV) DNA loads measured in plasma specimens by 1st WHO standard-normalized real-time PCR assays is uncertain.

Methods: This retrospective study compares the performance of Abbott m2000sp, Qiagen QIA Symphony SP, and KingFisher Flex platforms using plasma samples from allogeneic hematopoietic stem cell transplant recipients and plasma spiked with the CMV AD169 strain. The Abbott RealTime CMV PCR assay was used for CMV DNA quantitation.

Results: Maximum differences in CMV DNA loads quantified in plasma from 11 allo-HSCT and spiked plasma over a wide range of viral DNA concentrations (2.0–4.0 log₁₀ IU/ml) were ≤0.5 log₁₀ IU/ml.

Conclusions: The CMV DNA extraction efficiency of the platforms evaluated varies. The impact of these variations on CMV DNA loads quantified in plasma may not be clinically relevant.

© 2024 Sociedad Española de Enfermedades Infecciosas y Microbiología Clínica. Published by Elsevier España, S.L.U. All rights are reserved, including those for text and data mining, AI training, and similar technologies.

Impacto de plataformas automatizadas de extracción de ácidos nucleicos en las cargas de ADN de Citomegalovirus en plasma cuantificadas mediante PCR en tiempo real normalizada según el primer estándar internacional de la OMS

RESUMEN

Palabras clave:

Citomegalovirus (CMV)

Carga de ADN del CMV

Plasma

Extracción de ADN del CMV

Plataformas de extracción de ácidos nucleicos

Cuantificación de ADN del CMV

Introducción: Se desconoce si el uso de distintas plataformas de extracción de ácidos nucleicos afecta la magnitud de las cargas de ADN de citomegalovirus (CMV) cuantificadas mediante PCR en tiempo real normalizadas al primer estándar de la OMS.

Métodos: Comparamos retrospectivamente las plataformas Abbott m2000sp, Qiagen QIA Symphony SP y KingFisher Flex utilizando muestras de plasma de receptores de trasplante alogénico hematopoyético (alo-TPH) y plasma inoculado con la cepa CMV AD169. Las cargas virales se cuantificaron mediante el ensayo Abbott RealTime CMV PCR.

Resultados: Las diferencias máximas en las cargas cuantificadas en plasma de 10 alo-TPH y plasma inoculado, en un rango amplio de concentraciones (2,0 a 4,0 log₁₀ UI/ml) fueron ≤0,5 log₁₀ UI/ml.

Conclusiones: La eficiencia de extracción de ADN de CMV de las plataformas analizadas varía; sin embargo, el impacto de estas variaciones en las cargas de ADN del CMV cuantificadas en plasma podría no ser clínicamente relevante.

© 2024 Sociedad Española de Enfermedades Infecciosas y Microbiología Clínica. Publicado por Elsevier España, S.L.U. Se reservan todos los derechos, incluidos los de minería de texto y datos, entrenamiento de IA y tecnologías similares.

* Corresponding author.

E-mail address: david.navarro@uv.es (D. Navarro).

Introduction

Monitoring Cytomegalovirus (CMV) DNA load in blood (plasma or whole blood) is a cornerstone in the management of CMV infection in transplant recipients.¹ Normalization of CMV DNA loads, quantified by different commercially available nucleic acid amplification testing (NAAT) platforms, to the 1st WHO international standard² has led to improved agreement in CMV DNA load values across assays; yet, variability persists.³ The size of the PCR amplicon generated largely explains such variability^{4,5} as CMV DNA present in blood, most notably in plasma, is highly fragmented^{6–8}; nevertheless, other factors may also contribute. In this sense, the efficiency of nucleic acid extraction platforms was shown to have a sizeable impact on the magnitude of CMV DNA loads measured by different real-time PCRs prior to their normalization to the 1st WHO international standard.^{9,10} Moreover, by using small-sized synthetic oligonucleotides and a droplet digital PCR assay for CMV DNA quantitation, Cook et al.¹¹ showed that nucleic acid yields vary widely across extraction platforms, particularly for small DNA fragments (≤ 100 bp). Refined magnetic separation-based rapid nucleic acid extraction technology has been increasingly incorporated, both in automated nucleic acid extraction systems and high-throughput platforms, combining nucleic acid extraction and NAAT.¹² There is scarce information¹³ regarding how the performance of recent-generation nucleic acid extraction platforms impacts the magnitude of CMV DNA loads measured in plasma specimens by 1st WHO standard-normalized real-time PCR assays. Here, we addressed this issue by measuring CMV DNA loads in plasma from allogeneic hematopoietic stem cell transplant recipients (allo-HCT) after nucleic acid extraction in three widely used magnetic separation-based nucleic acid extraction platforms currently in use in our laboratory.

Material and methods

A total of 11 cryopreserved (-80°C for less than one year) plasma specimens from 11 allo-HCT patients who developed CMV DNAemia were retrieved for the analyses detailed below. CMV DNA load in these specimens had been routinely quantified by the Alinity m CMV assay (Abbott Molecular Inc., Des Plaines, IL, USA). For some experiments, plasma specimens testing negative for CMV DNA by the Alinity m CMV assay were spiked with different concentrations of the CMV AD169 strain. The following nucleic acid extraction platforms were used according to the different manufacturer's instructions: (i) Abbott m2000sp mSample Preparation System DNA; (ii) Qiagen QIAAsymphony SP DSP Virus/Pathogen Midi Kit (Hilden, Germany); (iii) KingFisher Flex MagMAX™ Pathogen RNA/DNA Kit (Waltham, Massachusetts, USA). Relevant characteristics of these extraction platforms are shown in [Supplementary Table 1](#). The RealTime CMV PCR (Abbott Molecular), a dual target (UL34 and UL80.5) PCR, with a limit of detection and quantification of 31.4 IU/ml (95% CI) was employed for CMV DNA quantitation. CMV DNA loads used for the analyses described below were normalized to the input and elution volumes of each extraction platform as follows: final (corrected) CMV DNA load in IU/ml: measured CMV DNA load in IU/ml $\times$ correction factor which is derived from: elution volume/PCR mix volume $\times$ units volume/input volume. This only applies to the KingFisher Flex platform. Means and coefficients of variation (CV%) of $\log_{10}$ transformed values, calculated using Excel 2021 (version 18.0) software, are reported throughout the study. The current study was approved by the Ethics Committee of the Hospital Clínico Universitario-INCLIVA (November 2022). The requirement for informed consent was waived by the Ethics Committee.

Results

A scheme of the study design is shown in [Supplementary Figure 1](#). The study initial experiments were performed using a single plasma specimen with a high CMV DNA load, as measured by the Alinity m CMV assay (83,718 IU/ml), which was conveniently diluted in pooled CMV DNA-negative plasma specimens, yielding the following CMV DNA concentrations: 2.0, 2.7, 3.0, 3.7 and 4.0 $\log_{10}$ IU/ml. Nucleic acid extraction was carried out in the three platforms on three consecutive days; nucleic acid eluates were then run in singlets on the m2000rt platform (a total of 45 tests). As shown in [Table 1](#), both the QIAAsymphony SP and KingFisher Flex, which performed similarly across all CMV DNA concentrations tested (maximum difference, 0.1 $\log_{10}$ IU/ml), were slightly more efficient than the m2000sp platform, according to the magnitude of CMV DNA loads measured. The maximum difference between QIAAsymphony SP and m2000sp was 0.40 $\log_{10}$ (at CMV DNA concentrations ≤ 2.7 $\log_{10}$ IU/ml), and between KingFisher Flex and m2000sp, it was 0.50 $\log_{10}$ (at CMV DNA concentrations of 2.7 $\log_{10}$ IU/ml).

To further assess the impact of the nucleic acid extraction yield of each platform on CMV DNA load values, 10 plasma specimens from 10 allo-HCT recipients with CMV DNA loads ranging from 13,492 to 195,872 IU/ml, as quantified by the Alinity m CMV assay, were serially diluted in pooled CMV DNA-negative plasma to achieve nominal viral DNA concentrations of 2.0, 2.7, 3.0, and 4.0 IU/ml. Plasma specimens were processed in the three platforms (once) and run on the m2000rt PCR system (in singlets) (a total of 120 tests). As shown in [Table 2](#), and [Supplementary Figure 2](#), the QIAAsymphony SP platform was slightly more efficient than the other two systems across all CMV DNA concentrations tested, although maximum differences in CMV DNA loads were ≤ 0.25 $\log_{10}$ IU/ml. Finally, we spiked pooled CMV DNA-negative plasma with increasing concentrations (2.0, 2.7, 3.0, 3.7 and 4.0 IU/ml) of a whole virus stock (CMV strain AD169), whose CMV DNA content had been quantified by the Alinity m CMV assay. Plasma specimens were processed in the three platforms on three consecutive days and then run in singlets on the m2000rt system ($n = 36$ tests). As shown in [Supplementary Table 2](#), overall, maximum differences across platforms did not exceed 0.5 $\log_{10}$ IU/ml, with the QIAAsymphony SP platform performing discretely better at some but not all CMV DNA concentrations.

Discussion

CMV DNA in plasma from transplant recipients is mostly unprotected and highly fragmented,^{6–8} with fragments ≤ 100 bp representing the largest fraction of viral DNA content.⁸ The efficiency of different commercially available extraction platforms to extract CMV DNA fragments of 50 or 100 bp was shown to vary widely across systems¹¹; nevertheless, in this study, specimens spiked with synthetic oligonucleotides analyzed via droplet digital PCR (ddPCR) for CMV DNA quantitation were used in the experiments. ddPCR is based on the isolated amplification of thousands of individual DNA molecules simultaneously, with each molecule compartmentalized in a droplet. The presence of amplified product in each droplet is indicated by a fluorescent signal, and the proportion of positive droplets allows the precise quantification of a given sequence in the absence of quantitation standards.¹⁴ Here, we examined whether three magnetic separation-based automated nucleic acid extraction platforms used for clinical plasma specimens significantly impacted CMV DNA loads, as measured by a 1st WHO standard-normalized real-time PCR assay. Two major observations were made. First, there were minimal differences, although sizeable, in the performance of the three extraction platforms when

Table 1
Cytomegalovirus DNA loads quantified by the RealTime CMV PCR assay on a single plasma specimen following nucleic acid extraction using different platforms carried out in three consecutive days.

Nominal CMV DNA load value in log ₁₀ IU/ml ^a	Mean CMV DNA load in log ₁₀ IU/ml (CV%) ^b		
	m2000sp	QIAasymphony SP	KingFisher Flex
2.0	2 (1.2)	2.4 (1.4)	2.3 (1.1)
2.7	2.5 (1.8)	2.9 (1.2)	3 (1.2)
3.0	3 (1.2)	3.2 (1.2)	3.3 (0.4)
3.7	3.7 (1)	3.8 (1.5)	3.9 (1.5)
4.0	3.9 (1.1)	4.1 (1.5)	4.1 (1.6)

CMV: Cytomegalovirus; CV: coefficient of variation; IU: International Units.
^a The CMV DNA load in the original specimen used for dilutions was measured by the Alinity m CMV assay.
^b Nucleic acid extraction was carried out in the three platforms on three consecutive days (in singlets).

Table 2
Cytomegalovirus DNA loads quantified by the RealTime CMV PCR assay in plasma from 10 allogeneic hematopoietic stem cell transplant recipients following nucleic acid extraction using different platforms.

Nominal CMV DNA load value in log ₁₀ IU/ml ^a	Mean CMV DNA load in log ₁₀ IU/ml (CV%) ^b		
	m2000sp	QIAasymphony SP	KingFisher Flex
2.0	2.2 (1.71)	2.4 (1.7)	2.3 (1.7)
2.7	2.8 (1.6)	3 (1.5)	2.7 (1.5)
3.0	3 (1.7)	3.3 (1.9)	3.2 (2.2)
4.0	3.8 (1.7)	4 (1.8)	3.8 (1.8)

CMV: Cytomegalovirus; CV: coefficient of variation; IU: International Units.
^a The CMV DNA load in the original specimen used for dilutions was measured by the Alinity m CMV assay.
^b Nucleic acid extraction was carried out in the three platforms on the same day in singlets.

clinical plasma specimens from allo-HCT recipients were tested; the QIAasymphony SP platform was slightly more efficient. Differences in CMV DNA loads measured following extraction with the three platforms were overall ≤ 0.5 log₁₀ IU/ml, and were more pronounced at low CMV DNA concentrations (≤ 2.7 log₁₀ IU/ml). This observation should be taken into consideration when plasma specimens need to be pre-diluted (i.e. in neonates) to achieve the minimum volume required for extraction according to the respective manufacturer. This figure aligns with previously published data. In effect, Kim et al.¹³ found a mean difference of -0.32 log₁₀ copies/ml between the QIAasymphony RGQ and QIAcube systems using the Artus CMV QS RGQ and RG assays for CMV DNA quantification. In turn, Bravo et al.⁹ reported CMV DNA loads quantified by the RealTime CMV PCR assay to differ by ≤ 0.4 log₁₀ copies/ml following extraction with the m2000sp system, the High Pure viral nucleic acid kit on the COBAS AmpliPrep system (Roche Diagnostics, Mannheim, Germany), and the EZ1 Virus 2.0 kit (Qiagen, Valencia, CA) on the BioRobot EZ1. In this context, a difference of <0.5 log₁₀ IU/ml across CMV NAAT assays is not considered clinically relevant, although this extent is debatable. Second, all extraction platforms evaluated appeared to perform comparably with clinical plasma specimens (highly fragmented CMV DNA) and plasma spiked with whole virus (CMV DNA mostly protected and thus non-fragmented). The main limitation of the current study is the scarce number of clinical specimens tested. In summary, our data indicated that, although the CMV DNA extraction efficiency of commercially available platforms varies, the potential impact of these variations on CMV DNA loads quantified in clinical plasma specimens by 1st WHO standard-normalized real-time PCR assays may be clinically irrelevant. Naturally, our assumption cannot be extended to extraction platforms and real-time PCR assays other than those evaluated herein.

Funding

This research did not receive any specific grant from funding agencies in the public, commercial, or not-for-profit sectors.

Authors’ contributions

Ángela Sánchez-Simarro, Eliseo Albert, Paula Michelena, and Estela Giménez: Methodology, investigation, formal analysis, data curation, and writing review & editing. David Navarro: Conceptualization, investigation, formal analysis, and writing the original draft.

Competing interests

The authors declare that they have no known competing financial interests or personal relationships that could have appeared to influence the work reported in this paper.

Acknowledgments

Ángela Sánchez-Simarro (PFIS Contract; FI22/00338) holds a contract funded by the Carlos III Health Institute. Eliseo Albert (Juan Rodés Contract; JR20/00011) holds a contract funded by the Carlos III Health Institute (co-financed by the European Regional Development Fund, ERDF/FEDER).

Appendix A. Supplementary data

Supplementary data associated with this article can be found, in the online version, at [doi:10.1016/j.eimc.2024.07.008](https://doi.org/10.1016/j.eimc.2024.07.008).

References

1. Ljungman P, de la Camara R, Robin C, Crottiolo R, Einsele H, Hill JA, et al. Guidelines for the management of cytomegalovirus infection in patients with haematological malignancies and after stem cell transplantation from the 2017 European Conference on Infections in Leukaemia (ECIL 7). *Lancet Infect Dis*. 2019;19:e260–72.
2. Fryer JF, Heath AB, Minor PD. A collaborative study to establish the 1st WHO International Standard for human cytomegalovirus for nucleic acid amplification technology. *Biologicals*. 2016;44:242–51.
3. Hayden RT, Sun Y, Tang L, Procop GW, Hillyard DR, Pinsky BA, et al. Progress in quantitative viral load testing: variability and impact of the WHO quantitative international standards. *J Clin Microbiol*. 2017;55:423–30.

4. Preiksaitis JK, Hayden RT, Tong Y, Pang XL, Fryer JF, Heath AB, et al. Are we there yet? Impact of the first international standard for Cytomegalovirus DNA on the harmonization of results reported on plasma samples. *Clin Infect Dis*. 2016;63:583–9.
5. Naegele K, Lautenschlager I, Gosert R, Loginov R, Bir K, Helanterä I, et al. Cytomegalovirus sequence variability, amplicon length, and DNase-sensitive non-encapsidated genomes are obstacles to standardization and commutability of plasma viral load results. *J Clin Virol*. 2018;104:39–47.
6. Boom R, Sol CJ, Schuurman T, Van Breda A, Weel JF, Beld M, et al. Human Cytomegalovirus DNA in plasma and serum specimens of renal transplant recipients is highly fragmented. *J Clin Microbiol*. 2002;40:4105–13.
7. Tong Y, Pang XL, Mabilangan C, Preiksaitis JK. Determination of the biological form of human Cytomegalovirus DNA in the plasma of solid-organ transplant recipients. *J Infect Dis*. 2017;215:1094–101.
8. Peddu V, Bradley BT, Casto AM, Shree R, Colbert BG, Xie H, et al. High-resolution profiling of human Cytomegalovirus cell-free DNA in human plasma highlights its exceptionally fragmented nature. *Sci Rep*. 2020;10:373.
9. Bravo D, Clari MÁ, Costa E, Muñoz-Cobo B, Solano C, José Remigia M, et al. Comparative evaluation of three automated systems for DNA extraction in conjunction with three commercially available real-time PCR assays for quantitation of plasma Cytomegalovirus DNAemia in allogeneic stem cell transplant recipients. *J Clin Microbiol*. 2011;49:2899–904.
10. Verheyen J, Kaiser R, Bozic M, Timmen-Wego M, Maier BK, Kessler HH. Extraction of viral nucleic acids: comparison of five automated nucleic acid extraction platforms. *J Clin Virol*. 2012;54:255–9.
11. Cook L, Starr K, Boonyaratanakornkit J, Hayden R, Sam SS, Caliendo AM. Does size matter? Comparison of extraction yields for different-sized DNA fragments by seven different routine and four new circulating cell-free extraction methods. *J Clin Microbiol*. 2018;56, e01061–18.
12. Li Y, Liu S, Wang Y, Wang Y, Li S, He N, et al. Research on a magnetic separation-based rapid nucleic acid extraction system and its detection applications. *Biosensors (Basel)*. 2023;13:903.
13. Kim H, Hur M, Kim JY, Moon HW, Yun YM, Cho HC. Automated nucleic acid extraction systems for detecting Cytomegalovirus and Epstein-Barr virus using real-time PCR: a comparison study between the QIAAsymphony RGQ and QIAcube systems. *Ann Lab Med*. 2017;37:129–36.
14. Vossen RH, White SJ. Quantitative DNA analysis using droplet digital PCR. *Methods Mol Biol*. 2017;1492:167–77.