

Routine HIV screening among blood donors in Buenos Aires (Argentina): Results from six years' experience and report of a single window-period donation

Silvina Alejandra Gendler and María Susana Pascuccio

Hemotherapy and Immunohematology Unit. Diagnosis and Treatment Department. Juan A. Fernández Hospital. Government of Buenos Aires City. Argentina.

BACKGROUND. Blood donor HIV antibody detection has been mandatory in Argentina since 1991, and p24 antigen screening was recommended in 1997.

METHODS. A total of 30,132 consecutive donations were screened. Repeatedly reactive samples were tested by another screening test and/or by Western blot (WB) for HIV Ab, or by a neutralization assay for p24 Ag.

RESULTS. Among the total, 0.3623% of samples were repeatedly reactive and 0.2084% were true HIV-infected donors. Only one donor tested nonreactive for HIV Ab, repeatedly reactive for p24 Ag, positive by neutralization assay, and seroconverted later. Samples with a signal-to-cutoff (S/CO) ratio ≥ 3.00 on routine HIV Ab testing were 100.0% positive by WB and/or repeatedly reactive by the second test. In samples with a S/CO ratio < 3.00 , 11.1% were positive by WB and/or the majority were nonreactive by the second test. Among HIV-infected donors, 89.5% possessed risk factors (which had been denied previously), 56.5% were repeatedly reactive by other screening procedures and 88.6% were coinfecting with other blood-transmissible viruses.

CONCLUSIONS. When the EIA S/CO ratio is ≥ 3.00 , WB can be replaced by a second screening test. The pre-donation questionnaire should be improved to detect risk behavior in prospective donors. There was a high association between HIV and other blood-transmissible viruses.

Key words: HIV antibodies. HIV core protein p24. Blood donors. Risk factors.

Pruebas de cribado sistemáticas para VIH entre donantes de sangre de Buenos Aires, Argentina: resultados de una experiencia de 6 años y descripción de una donación en período de ventana

ANTECEDENTES. En Argentina, la detección de anticuerpos anti-VIH (virus de la inmunodeficiencia humana) es

obligatorio desde 1991 y desde 1997 se recomienda el tamizaje del antígeno p24.

MÉTODOS. Se tamizaron un total de 30.132 donaciones sucesivas. A las muestras repetidamente reactivas se les realizó otra reacción de tamizaje y/o Western blot (WB), para anticuerpos anti-VIH, o ensayo de neutralización para el antígeno p24.

RESULTADOS. Se obtuvieron el 0,3623% de muestras repetidamente reactivas y el 0,2084% de donantes verdaderamente infectados con VIH. Sólo un donante resultó no reactivo para anti-VIH, repetidamente reactivo para antígeno p24 y positivo por neutralización con posterior seroconversión. Las muestras con razón S/CO $\geq 3,00$ en los ensayos de rutina para anti-VIH fueron 100,0% positivos para WB y/o repetidamente reactivos para una segunda reacción. Las muestras con razón S/CO $< 3,00$ fueron 11,1% positivas por WB y/o la mayoría resultó no reactiva para un segundo ensayo. El 89,5% de los donantes infectados con VIH poseían factores de riesgo negados previamente. El 56,5% de los donantes infectados con VIH fueron repetidamente reactivos para otras reacciones de tamizaje y el 88,6% de ellos presentó coinfección con otros virus transmisibles por sangre.

CONCLUSIONES. Cuando la razón del ensayo de enzimoimmunoanálisis $\geq 3,00$, el WB podría ser reemplazado por una segunda reacción de tamizaje. Es necesario mejorar el cuestionario predonación a fin de detectar comportamientos de riesgo entre los posibles donantes. Hubo una alta asociación entre VIH y otros virus transmisibles por sangre.

Palabras clave: Anticuerpos anti-VIH. Antígeno p24. Donantes de sangre. Factores de riesgo.

Introduction

In 1982, the *Morbidity and Mortality Weekly Report* from the Centers for Disease Control (CDC) provided information on possible transmission of acquired immunodeficiency syndrome (AIDS) by contaminated blood and blood products¹. Since then, public and private health institutions, blood banks and manufacturers of diagnostic tests have collaborated to develop a blood safety vigilance system to reduce transfusion-related human immunodeficiency virus (HIV) transmission.

Correspondence: Dra. S.A. Gendler.
Migueletes, 1127 2° B. 1126 Buenos Aires, Argentina.
E-mail: gendlersa@yahoo.com.ar

Manuscrito recibido el 6-11-2005; aceptado el 8-6-2006.

At the beginning of this effort, most HIV-contaminated blood units were discarded by surrogate marker screening². In 1985, the first specific HIV type 1 (HIV-1) antibody test was licensed and blood donors then underwent double screening: a) an extensive pre-donation questionnaire to determine donor eligibility and b) post-donation HIV-1 antibody testing. The most commonly used method for detecting HIV antibodies (HIV Ab) has been enzyme-linked immunosorbent assay (EIA). The performance of HIV screening assays has improved substantially over the years. First-generation HIV tests, based on HIV-1 whole-virus lysate, were able to detect only anti HIV-1 IgG. Subsequently, replacement of viral lysate antigens with recombinant proteins and synthetic peptides that detect anti HIV-1 and HIV-2 IgG (second-generation tests) increased the sensitivity of the assay, particularly in the early phase of seroconversion. Later, third generation HIV Ac testing, and HIV-1 p24 assays, which detect the p24 core antigen produced by the HIV-1 *gag* gene, were developed and introduced in blood safety vigilance. When the HIV-1 antigen (p24 protein) test is implemented together with HIV-1/2/O antibody screening, the window period is lowered to 16-17 days³. New combined antigen/antibody testing, the so-called fourth generation tests, are more sensitive than third generation antibody tests; nevertheless, their analytical sensitivity for p24 Ag (p24 Ag) is lower than that of single antigen tests⁴. Nucleic acid amplification technology can improve the effectiveness of HIV detection and shorten the potentially infectious window to 10-12 days³. However, this testing technology has an elevated cost and requires separate laboratory areas and highly trained personnel.

Detection of HIV Ab has been mandatory in Argentina since 1991, and the *Asociación Argentina de Hemoterapia e Inmunohematología* (AAHI; Argentine Society for Hemotherapy and Immunohematology) recommended routine screening for p24 Ag in 1997, when p24 Ag tests were commercially available in our country. Use of the p24 Ag assay was implemented on 4 September 1997 to ensure a safer blood supply for our hospital patients who needed transfusions.

The aim of this retrospective study is to describe the results obtained during double HIV screening (antibodies to HIV 1/2/O and p24 Ag) among our blood donors, and clinical correlations with these results, and to report the single case of HIV seroconversion detected over the time since screening has been practiced.

Methods

Blood donors

All blood donors at our blood bank are unpaid; most of them are occasional directed donors who voluntarily donate blood for the transfusion needs of a relative or friend (allogeneic donors). When patients are able to donate blood for themselves, autologous blood donations are also drawn. Prior to blood donation, all prospective donors receive an oral explanation about the donation process and utility of the transfusion, and are confidentially evaluated about clinical and infectious conditions according to the AAHI recommendations. Blood pressure, pulse, temperature, weight and hematocrit are checked. Donors must complete a general health history questionnaire. A more detailed version of this questionnaire was implemented in January 2002. Donors are then asked to sign a written informed consent form allowing serologic screening and have the opportunity to ask ques-

tions. All blood donors can voluntarily choose to leave at any point of the process without donating, and have the chance to exclude themselves as blood donors during the time following the donation (self-deferral form, implemented in 2001). Donors with repeatedly reactive or positive serological test results are routinely recalled by mail citation. If they return, an extensive and more detailed questionnaire is given to detect risk behavior or a disease history that was denied or omitted during the pre-donation interview. Additional serological tests are performed when necessary. These individuals are referred to the Infectious Disease Service to be given information on follow-up diagnostic assessment and the available medical, preventive and psychosocial services.

Between 4 September 1997 and 30 September 2003, a total of 30,132 consecutive blood donations were completed and screened for transfusion-transmitted infections, including presence of HIV Ab and p24 Ag. Individual data that we report in this paper were obtained from the pre-donation questionnaire and the clinical files of the Infectious Disease Service.

Samples

All fresh serum specimens were immediately examined on the same day as collection by screening tests. Initially reactive samples were kept at 4 °C and retested in duplicate within the three following days. Repeatedly reactive specimens were frozen at -20 °C and thawed once prior to confirmatory testing.

HIV antibodies and HIV-1 antigen: dual testing algorithm

The routine blood donor testing algorithms for HIV Ab and HIV-1 Ag are shown in figure 1. Some additional considerations are mentioned:

- Samples from blood donors were always analyzed by two HIV screening methods (one for HIV Ab and the other for HIV-1 Ag). According to the 1997 revised UNAIDS-WHO recommendations⁵, HIV antibody testing strategy I was applied, and tests having the highest possible sensitivity were used to minimize the rate of false-negative results and to ensure transfusion safety. Because the fourth-generation HIV Ab tests we used have a low analytical sensitivity for p24 Ag, this assay was used only to detect HIV Ab.

- In accordance with CDC guidelines⁶ and the manufacturer's instructions, any sample with an initially reactive result was retested in duplicate. If the result for either duplicate test was reactive, the specimen was reported as repeatedly reactive and the blood unit was discarded.

- When confirmatory assays could not be performed, one or two additional screening tests were used, in keeping with proposed WHO HIV antibody testing strategies II or III⁵ for diagnostic purposes, before notifying any repeatedly reactive HIV result to a blood donor.

- Western blot (WB) interpretation was done according WHO criteria⁷. When a sample showed an inadequate band pattern to meet the recommended criteria for WB positive status, it was considered indeterminate⁸.

- Neutralization of viral antigen by specific antibody is an accepted confirmatory method for p24 Ag repeatedly reactive samples. This neutralization assay was performed promptly^{9,10} because of the potential deterioration of HIV antigen during sample storage.

Other serological screening tests

Besides anti HIV and p24 Ag testing, eight other screening tests were performed on each blood donor sample to detect anti HBc and HBsAg for hepatitis B, anti HCV, anti HTLV I-II, anti *Brucella abortus*, serologic markers for syphilis by treponemal or nontreponemal methods, and anti *Trypanosoma cruzi* by two different methods (mandatory) to detect Chagas' disease.

Statistical analysis

All statistical calculations were performed on a personal computer using Microsoft Excel software or the MEDCALC demo for Windows.




Figure 1. Routine blood donor testing algorithm for HIV Ab and HIV-1 Ag.

TABLE 1. Screening and confirmatory assays and analyzers used over the period assessed***

Trademark and manufacturer	Assay design
HIV Ab screening assays	
<i>SFD HIV 1/2 PA</i> from Sanofi Fujeribo, Japan (previously) and Bio-Rad, France (later)	Passive particle agglutination test using gelatin particles sensitized with recombinant HIV-1 and HIV-2 antigens
<i>Abbott Recombinant HIV1/HIV2 3rd Generation EIA</i> (and <i>plus</i> version)	EIA based on a "sandwich" principle that detects anti HIV-1/2 antibodies (3 rd generation)
<i>Genelavia® MIXT</i> from Sanofi Diagnostics Pasteur, France	microEIA based on an indirect principle that detects anti HIV-1/2 antibodies (2 nd generation)
<i>Vironostika® HIV MIXT</i> from Organon Teknika, The Netherlands	microEIA based on a "sandwich" principle that detects anti HIV-1/2 antibodies (3 rd generation)
<i>Vironostika® HIV Uni-Form II</i> from Organon Teknika, The Netherlands	microEIA based on a one-step "sandwich" principle anti HIV-1/2 antibodies (3 rd generation)
<i>Vironostika® HIV Uni-Form II plus O</i> from Organon Teknika (previously) and Biomerieux (currently), The Netherlands	
<i>ORTHO® HIV-1/HIV-2 Ab Capture ELISA Test System</i> from Ortho-Clinical Diagnostics, Inc., USA	microEIA based on a one-step "sandwich" principle that detects anti HIV-1/2/O antibodies and p24 Ag (3 rd generation)
<i>Murex HIV-1.2.0</i> from Murex Biotech Limited, UK	
<i>Vironostika® HIV Uni-Form II Ag/Ab</i> from Organon Teknika (previously) and Biomerieux (currently), The Netherlands***	microEIA based on a one-step "sandwich" principle that detects anti HIV-1/2/O antibodies and p24 Ag (4 th generation)
p24 Antigen screening assays	
<i>Vironostika® HIV-1 Antigen</i> from Organon Teknika (previously) and Biomerieux (currently), The Netherlands	
<i>Coulter HIV-1 p24 Antigen Assay</i> from Coulter Corporation, USA	microEIA based on "sandwich" principle for detection of the HIV-1 p24 core antigen
<i>Murex HIV Antigen Mab</i> from Murex Biotech Limited, UK	
Microplate analyzers	
<i>Tektime</i> from Organon Teknika (previously) and Biomerieux (currently), The Netherlands	microEIA analyzer
<i>Biomaster</i> from Biokit, Spain	microEIA analyzer
HIV Ab confirmatory tests	
<i>HIV-1 Western Blot Kit</i> from Organon Teknika Corporation, USA	Nitro-cellulose strips with electroblotted antigenic proteins from partially purified inactivated HIV-1 viral particles (whole cell viral lysate) to detect anti HIV-1 antibodies
<i>New Lav Blot I</i> from Sanofi Diagnostics Pasteur, France	
<i>HIV Blot 2.2</i> from Genelabs Diagnostics, Singapore	Nitro-cellulose strips with electroblotted antigenic proteins from partially purified inactivated HIV-1 viral particles and specific HIV-2 synthetic peptide to detect anti HIV-1/2 antibodies
p24 Antigen Neutralization test	
<i>Vironostika® HIV-1 Antigen Neutralization System</i> from Organon Teknika (previously) and Biomerieux (currently), The Netherlands	Pre-test treatment with specific antibodies that neutralize HIV-1 p24 Ag

*No chronological order was followed (some kits were used several times during different periods).

**All assays were performed in accordance with the manufactures' instructions.

***This assay was used only to detect HIV Ab. A test for single p24 Ag detection was performed at the same time.

Results are expressed as mean \pm SD and/or 95% CI, where applicable. When the number of samples was large enough, the test of hypothesis for two populations (TH) was used. Data were analyzed with the 2×2 chi-square test with Yates' correction, or by Fisher's exact test, where applicable. Significance was set at a *P*-value of less than .05. Specificity was calculated as the ratio between true-negative results and HIV non-infected donors (true-negative + false-positive donors)⁵.

Results

Among the 30,132 blood donor samples screened, 108 (0.3623%; 95% CI, 0.3611-0.3636) were repeatedly reac-

tive by HIV Ab and/or p24 Ag screening tests, mainly EIA (table 1). Donors were classified into three groups according to the screening and confirmatory test results, which are summarized in table 2a. Among the 108 HIV Ab and/or p24 Ag repeatedly reactive donors only 62 (57.41%) were confirmed as true-positive, which yields 0.2084% (62 of 30,132; 95% CI, 0.2072-0.2096) of true HIV-infected donors.

The percentage of true-positive results was significantly higher (*P* < 0.0001) in HIV Ab repeatedly reactive donors (groups A + B: 61 of 87 [70.11%]) as compared to p24 Ag repeatedly reactive donors (groups B + C: 5 of 25 [20.00%]).

TABLE 2. Distribution of blood donors according to HIV screening and confirmatory test reactivity

(2a) Number of blood donors for each condition				
Group	HIV Ab test reactivity	p24 Ag test reactivity	True-Positive ^a	Total
A	RR ^b	NR	57	83
B	RR	RR	4	4
C	NR ^c	RR	1	21
Total	–	–	62	108

(2b) S/CO ^d ratio for each condition (mean ± SD)				
Groups	RR at least for	True-Positive ^a	False-Positive ^e	P
A + B	HIV Ab	11.72 ± 4.28	1.36 ± 0.54	< 0.0001
B + C	p24 Ag	7.87 ± 7.71	1.75 ± 1.33	0.0017

(2c) Age and sex distribution for each condition (mean ± SD)				
	RR donors (groups A + B + C)			Donor population
	True-Positive ^a	False-Positive ^e	Total	Total
n	62	46	108	30,024
Age (years, mean ± SD)	33.8 ± 9.0 ^f	36.0 ± 11.4 ^f	34.5 ± 9.9 ^g	34.2 ± 11.2 ^g
Sex: male/female (% fem)	55/7 (11.29) ^h	34/12 (26.09) ^{h,i}	89/19 (17.59) ^j	1,226/419 (25.47) ^{i,j}

^aRR samples by screening tests that were confirmed positive by confirmatory tests.

^bRepeatedly reactive result (when EIA S/CO ratio of the sample is ≥ 1.0).

^cNonreactive result (when EIA S/CO ratio of the sample is < 1.0).

^dSignal-to-cutoff ratio (calculated by dividing the optical density value of the sample being tested by the optical density value of the assay cutoff for that run).

^eRR samples by screening tests that were not confirmed positive by confirmatory tests.

^fStatistically significant differences (33.8 ± 9.0 vs. 36.0 ± 11.3; *P* < .1).

^gNo statistically significant differences (34.2 ± 1.2 vs. 34.5 ± 9.9).

^hStatistically significant differences (11.29% vs. 26.09%; *P* < .05).

ⁱNo statistically significant differences (26.09% vs. 25.47%).

^jStatistically significant differences (17.59 vs. 25.47; *P* < .05).




Figure 2. Evolution of p24 Ag concentration and HIV Ab EIA S/CO from donation day (day 0).

Among our 21 donors who tested repeatedly reactive by p24 Ag EIA and nonreactive by HIV Ab testing (group C), only one donor was found to be positive by neutralization assay (4.76%). This donor was recalled and a serological and virological HIV follow-up was performed on the initial and two additional successive specimens in coordination with the Infectious Disease Service. It was not possible to establish the date of the sexual event that triggered the

infection. The donor showed evidence of seroconversion in a second sample 13 days after the donation (fig. 2 and table 3)¹¹.

It is known that the proportion of screening-test-reactive results that are positive by confirmatory tests increases as the screening-test-reactive average S/CO ratio increases. In our donor population, there were significant differences in the S/CO ratio (*P* < 0.0001) between EIA true-positive HIV Ab repeatedly reactive donors and false-positive donors (11.72 ± 4.28 vs. 1.36 ± 0.54) (groups A and B, table 2b). The same was true among p24 Ag repeatedly reactive donors by EIA (7.87 ± 7.71 vs. 1.75 ± 1.33; *P* = 0.0017) (groups B and C, table 2b). The proportion of HIV Ab repeatedly reactive donors by EIA that tested positive by WB was 11.1% for samples with an average S/CO ratio < 3.0 and 100.0% for those with an average S/CO ratio ≥ 3.00 (fig. 3). In addition, 31 of 85 (35.6%) HIV Ab repeatedly reactive samples were tested by a second screening test; another EIA, a microparticle enzyme immunoassay, or an agglutination test (table 4). All 13 samples with an S/CO ratio ≥ 3.0 (10.72 ± 3.84) on the first test were also reactive on the second test and positive by WB. However, the 18 samples that had an S/CO ratio < 3.0 (1.47 ± 0.58) with the routine screening test were nonreactive (15), reactive (1) or borderline (2) with the second test. All these donors were confirmed as HIV non-infected.

At the time of blood donation, all donors had denied high risk behaviors for HIV infection and, therefore, completed the donation process. Nevertheless, some of them were found to be repeatedly reactive for HIV Ab and/or p24 Ag, including one donor who self-deferred (he was

TABLE 3. Follow-up of the single p24 Ag-reactive patient who later seroconverted

	Sample N°		
	1	2	3
Days from donation day	0	13	21
p24 Ag EIA result (S/CO)	RR (16.80)	RR (4.98)	NR ^a (0.88)
p24 Ag quantification ^b (pg/mL)	312	83	26
p24 Ag neutralization assay result, by (a) commercial and/or (b) in house method	positive (both methods were performed)	positive (only (b) was performed)	^c
HIV Ab EIA result (S/CO)	NR ^a (0.34)	RR (3.39)	RR (2.79)
HIV WB result	^c	indeterminate	positive for HIV-1
HIV-1 RNA detection assay ^d	positive	^c	^c
Medical follow-up	no symptoms	no symptoms	no symptoms

^aNR: nonreactive result (when EIA S/CO ratio of the sample is < 1.0).

^bp24 Ag was quantified using the same screening EIA test (a calibration curve was performed according to the manufacturer's instructions).

^cNot done.

^dby reverse transcription followed by polymerase chain reaction (RT-PCR), Nuclisens NASBA (Organon Teknika, The Netherlands).

found to be false-positive for HIV Ab). All the repeatedly reactive donors (n = 108) were requested by mail to return.

The true-positive donors (n = 62) were divided into three groups: 1) 17 donors (27.4%) who did not come to the appointment, six of whom had given a false address; 2) 21 donors (33.9%) for whom additional information could not be obtained because they decided to receive medical care outside of our hospital; and 3) 24 donors (38.7%) who returned and were referred to the Infectious Disease Service. After extensive lifestyle questioning, it was found that 89.5% of the HIV-infected donors who returned possessed risk factors for retroviral infection, including unprotected heterosexual (47.4%), homosexual (15.9%) and bisexual (15.9%) activity and intravenous drug use (10.5%). A high proportion of infected donors (70.0%) did not know their partner's HIV serostatus and only one infected donor admitted sexual activity with a risk partner. The seroconverting donor, the only true-positive donor from group C (p24 Ag repeatedly reactive by EIA only),




Figure 3. Distribution of S/CO ratios from HIV Ac EIA screening versus HIV infection status.

TABLE 4. Results obtained with a second HIV Ab screening test in some of the repeatedly reactive samples tested with the routine method

Samples confirmed as non HIV positive			Samples confirmed as HIV positive		
N	S/CO ratio routine EIA	Result for the second test	N	S/CO ratio routine EIA	Result for the second test
1	1.0	nonreactive	1	4.9	all reactive
2	1.1	borderline	2	6.5	
3	1.1	nonreactive	3	7.0	
4	1.1	reactive	4	8.2	
5	1.1	nonreactive	5	8.8	
6	1.1	nonreactive	6	9.6	
7	1.1	nonreactive	7	10.1	
8	1.1	nonreactive	8	13.0	
9	1.1	nonreactive	9	13.0	
10	1.3	nonreactive	10	13.0	
11	1.3	nonreactive	11	13.0	
12	1.3	nonreactive	12	13.0	
13	1.3	reactive	13	19.3	
14	1.7	nonreactive	mean	10.72	4.9-19.3
15	2.0	nonreactive	SD	3.84	
16	2.4	nonreactive	range		
17	2.5	nonreactive			
18	2.9	nonreactive			
mean	1.47				
SD	0.58				
range	1.0-2.9				




Figure 4. Additional reactivity from true (A) and false-positive (B) HIV Ab and/or p24 Ag donors.

was a 29-year-old man who later admitted heterosexual promiscuity.

No statistically significant differences in average age were observed between HIV Ab and p24 Ag non-reactive donors and HIV Ab and/or p24 Ag repeatedly reactive donors (34.2 ± 11.2 vs. 34.5 ± 9.9 years; table 2c). However, between true-positive HIV Ab and/or p24 Ag donors and false-positive donors (33.8 ± 9.0 vs. 36.0 ± 11.3 , a significant (although low) difference ($P < .1$) was found by TH.

A significantly smaller percentage of female donors was seen in the HIV Ab and/or p24 Ag repeatedly reactive group in comparison with non-reactive donors (17.59% vs. 25.47%) ($P < .05$). Similarly, the proportion of women was significantly lower ($P < .05$) in the true-positive groups as compared to the false-positive group (11.29% vs. 26.09%). The proportion of women in the false-positive group was similar to that of the negative group (26.09% vs. 25.47%, non-significant).

Among the total repeatedly reactive donor population, only 7.7% were reactive for two or more serological markers of infection (coinfections). However, among the 62 true-positive HIV donors, 35 (56.5%) tested repeatedly reactive for other screening reactions (3 for syphilis; 1 for Chagas' disease; 6 for anti HBc; 2 for anti HBc and HBsAg; 7 for anti HBc and anti HCV; 6 for anti HBc, anti HCV and anti HTLV I-II; 8 for anti HCV and 2 for anti HTLV I-II) and 43.5% were reactive only for HIV tests (fig. 4, part A). This proportion of coinfections was significantly higher (TH, $P < .01$) for HIV true-positive donors than for the overall reactive donors. Of 46 false-positive HIV donors, 8.7% were found to be reactive for two or more serological markers of infection (proportion similar to that of the overall infected donor population) whereas 91.3% showed reactivity (false) only for HIV serological markers (fig. 4, part B). Thirty-one of 35 (88.6%) HIV true-positive donors who presented coinfections were infected with other blood-transmissible viruses (HBV and/or HCV and/or HTLV).

Discussion

Our total rate of HIV repeatedly reactive donors (0.3623%), including HIV Ab and/or p24 Ag reactive

donors, was similar to values reported by other centers in our city¹² and Brazilian cities^{13,14}, but was higher than values reported by other blood banks in Argentina and the rest of Latin America¹⁵⁻¹⁹. The rate of confirmed HIV-infected donors (0.2084%) was intermediate among other public and private Latin American blood donation centers^{13,17,20,21}, but much higher than rates from other countries with a repeat donor population, such as the USA and France²².

Our data (80% of false-positive results among p24 Ag repeatedly reactive donors) are consistent with the concept that that most repeatedly reactive p24 Ag screening results in the blood donor population are expected to be false-positive or invalid^{9,18}. In contrast, most of the repeatedly reactive results in the HIV Ab assays were confirmed as positive (70.11%). These findings yield a specificity of 99.93% for p24 Ag and 99.91% for HIV Ab (the kits used report 100% sensitivity; hence all nonreactive samples were considered true negative). The specificity values obtained coincide with the values provided in the kit inserts and are compliant with the WHO minimum standards⁵. Although these specificity values for screening methods are acceptable, we think it is desirable to improve this parameter in order to minimize unnecessary donor deferrals.

To confirm repeatedly reactive samples for HIV Ab screening tests, supplemental or confirmatory tests, such as WB, should be performed. One-third of our HIV Ab repeatedly reactive donors were false positive; thus, it is very important to confirm HIV infection by additional tests before reporting reactive results to the donor to avoid the psychological distress associated with a false-positive HIV result. Other blood banks have expressed similar considerations¹³. Given the fact that confirmatory testing (such as WB) may sometimes be unavailable because of its high cost, and to minimize the number of specimens that require additional testing, we believe that in cases with an S/CO ratio ≥ 3.00 (or using another specific S/CO ratio established with a larger number of samples), WB can be replaced by a second screening test, in keeping with WHO recommendations⁵. If a sample is reactive on both these tests, it is considered HIV Ab-positive without wasting time and money performing WB testing. However, samples with low S/CO ratio (< 3.0) must be tested by WB to

define the true HIV status, as is recommended in the CDC algorithm for HCV²³.

Only one case of HIV seroconversion was detected among the 30,132 tested donors; that is, 3.3 cases per 100,000 donations. It has been estimated that one p24 Ag reactive/HIV Ab nonreactive donor will be present in every 88,000-175,000 donations in the Latin American replacement blood donor population¹⁸; therefore, it is probable that we will have to wait several years to detect a second seroconversion case in our donor population. p24 Ag is usually detectable a mean of six days earlier than HIV Ab²⁴. Based on the data in figure 2 it can be inferred that the donor might have seroconverted between the 4th and 6th day after donation, when p24 Ag was detected. Other blood banks in Argentina have reported cases of p24 Ag true-positive donors who later seroconverted^{25,26}. These results support the policy of continuing with p24 Ag testing among blood donors, despite the fact that this antigen is only detectable during a short time.

Among the group of true-positive HIV donors, a large number did not return after the mail citation or lied about their address. The low percentage of positive HIV donors that returned is similar to that observed by others¹⁴. It may be that the donors who did not return knew about or suspected their positive serological status and, nevertheless, came to the blood bank because of pressure by their family or work partners. Another possibility is that these individuals used the donation opportunity to obtain free HIV testing, which in our setting is only prescribed after a medical interview (these individuals may not have wanted a physician consultation)²⁷. Most of our infected donors were young men practicing high-risk behavior (unprotected sex activity and intravenous drug use), as has been reported by other blood banks and infectious disease centers^{17,20,28,29}. It is known that donors generally do not reveal their risk factors during the interview³⁰.

There was no clear demographic profile to distinguish between repeatedly reactive or infected donors and non-reactive ones. Furthermore, there were no significant differences in the mean age between nonreactive and repeatedly reactive donors for both HIV Ab and p24 Ag testing; however, among repeatedly reactive donors there was a small age difference ($P < .1$) between true- (younger) and false-positive donors.

A high association was found between HIV infection and other diseases, particularly blood-transmitted infections (coinfection with other blood-transmissible viruses was common). Similar coinfection patterns have been reported by other blood banks^{17,19-21} and health centers^{28,29}. In contrast, coinfection was rare among our HIV-negative donors, also in keeping with data from other blood banks¹⁹. It is clear that a HIV Ab and p24 Ag screening by improved methods and kits, more rigorous donor selection, and an increase in the proportion of voluntary and repeat donors versus replacement donors is necessary to improve transfusion safety. However, there still remains a small but significant transfusion risk for HIV and other viral agents. To reduce this residual risk, progressive introduction of nucleic acid amplification technology is necessary in blood screening³¹. Considering that RNA testing at a sensitivity of 50 copies/mL would detect HIV infection approximately 7 days before p24 Ag testing³², implementation of nucleic acid amplification for HIV

screening in our blood bank should be done as soon as possible to minimize the serological diagnostic window of recent HIV infection and reduce the residual transfusion-transmissible infection risk. Several publications have reported the benefits of this new technology in blood donor screening, in addition to serological testing^{31,33-37}. However, before introducing these techniques, associated issues must be resolved, such as the higher HIV screening cost per blood unit, and the need for separate laboratory areas and highly trained personnel for this purpose. A clear public health policy in favor of implementing nucleic acid amplification technology in blood banks would be an important advance to achieve this goal.

Acknowledgements

The authors thank Dr. Luisa Carnevali (chief of the Hemotherapy and Immunohematology Unit, Juan A. Fernández Hospital, Government of Buenos Aires City) for her kind cooperation and Dr. Pedro Cahn, Dr. Héctor Pérez and Lic. Florencia Ameal (Infectious Disease Unit, Juan A. Fernández Hospital, Buenos Aires) for clinical data collection and critical review of our manuscript, as well as Dr. Beatriz Livellara (Molecular Biology Laboratory, Italian Hospital, Buenos Aires) who performed the molecular study of our single HIV pre-seroconversion donor.

Contributions

The authors contributed equally during the preparation of this paper and the order of appearance of the names does not reflect differences in amount of work put into the writing of manuscript.

Disclosures

Conflict of interest: none.

Redundant publications: no substantial overlapping with previous papers.

References

1. Centre for Disease Control: Update on acquired immunodeficiency syndrome (AIDS) among patients with haemophilia A. *MMWR*. 1982;31:644-51.
2. Leitman SF, Klein HG, Melpolder JJ, et al. Clinical implications of positive tests for antibodies to human immunodeficiency virus type 1 in asymptomatic blood donors. *N Engl J Med*. 1989;321:917-24.
3. Schreiber GB, Busch MP, Korelitz JJ. The risk of transfusion-transmitted viral infections. The Retrovirus Epidemiology Donors Study. *N Engl J Med*. 1996;334:1685-90.
4. Polywka S, Feldner J, Duttman H, Laufs R. Diagnostic evaluation of a new combined HIV p24 antigen and anti-HIV-1/2/O screening assay. *Clin Lab*. 2001;47:351-6.
5. WHO Weekly Epidemiological Record: Revised recommendations for the selection and use of HIV antibody tests. 1997;72:81.
6. Centre for Disease Control: Revised Guidelines for HIV Counselling, Testing and Referral. *MMWR*. 2001;50(Nº RR-19).
7. WHO Weekly Epidemiological Record: Proposed WHO criteria for interpreting results from Western blot assays for HIV-1, HIV-2 and HTLV-I/HTLV-II. 1990;65:281.
8. Centre for Disease Control: Interpretation and use of Western blot assay for serodiagnosis of human immunodeficiency virus type 1 infections. *MMWR*. 1989;38 Suppl 7:1-7.
9. Food and Drug Administration, Department of Health and Human Services memorandum: Recommendations for Donor Screening with a Licensed Test for HIV-1 Antigen. 1995.
10. Gendler SA, Pascuccio MS, Carnevali L. Evaluación de la neutralización del antígeno p24 de HIV-1. *Rev Arg Transf*. 2001;XXVII Supl:32.
11. Gendler SA, Pascuccio MS, Carnevali L. Importancia de la determinación de HIV 1 Ag en el Banco de Sangre. A propósito de un caso. *Rev Arg Transf*. 2001;XXVII Supl:27.
12. Rey J, Tomeo A, Insua M, Santamarina D, Zárate N, Arrayago A, Ceriani S, et al. Prevalencia de anti-HIV, anti-HCV y HBsAg en dadores de sangre

- y pacientes prequirúrgicos en un Hospital de alta complejidad. *Rev Arg Transf.* 2003;XXIX Supl:28.
13. Ditzel DRC, Merolli R, Baldanzi GR, Almeida PTR. Teste para HIV, avaliar ensaios imuno enzimáticos (E.I.E.) e Western Blot (W.B.). *Boletim da Sociedade Brasileira de Hematologia e Hemoterapia.* 1998;XX Supl:72.
14. Otani MM, Salles NA, Barreto AM, Barreto CC, Chamone DF, Sabino EC. Evaluation of the concomitant use of two different EIA tests for HIV screening in Blood Banks. *Rev Panam Salud Publica.* 2003;13:172.
15. García MD, Talanda C, Pugliese P, Podzun I, Schmee E. Perfil serológico e imunohematológico de donantes de sangre. *Rev Arg Transf.* 2003;XXIX Supl:19.
16. Pivetta M, Chialina S, Nicolich V, Raillon M, Solís E. Estudio multicéntrico de seroprevalencia de enfermedades transmisibles por transfusión. *Rev Arg Transf.* 2003;XXIX Supl:30.
17. Pugliese MA, Payero MT, Ladino S, Rodríguez E, Fernández RJ. Evaluación de los marcadores serológicos en Banco de Sangre. *Rev Arg Transf.* 2001; XVII Supl:35.
18. Wendel S, Fachini RM, Oliveira LC, Levi JE, Lima LP, Hamada GS, et al. Blood Donor screening with p24 Antigen in Brazilian Blood Banks. *Boletim da Sociedade Brasileira de Hematologia e Hemoterapia.* 1998;XX Supl:20.
19. García CA, Agassis ACR, Marques Pereira JP, Soares IP, Farinon J, Balsan AM, et al. Anticorpo anti HIV I/II e Sua Associação com Outros Marcadores Sorológicos em Doadores de Sangue do Hospital de Clínicas de Porto Alegre: Atualização. *Boletim da Sociedade Brasileira de Hematologia e Hemoterapia.* 1998;XX Supl:71.
20. Blejer JL, Natalutti GR, García MM, Saguier MC, Salamone HJ. Infecciones por retrovirus en donantes de sangre. *Rev Arg Transf.* 2001;XXVII Supl:25.
21. Blejer JL, Chiera AO, Saguier MC, Salamone HJ. HIV P24 Antigen Screening in Two Blood Banks in Argentina. *Transfusion.* 2004;44 Supl:93A.
22. Le Corfec E, Le Pont F, Tuckwell HC, Rouzioux C, Costagliola D. Direct HIV testing in blood donations: Variation of the yield with detection threshold and pool size. *Transfusion.* 1999;39:1141.
23. Centre for Disease Control: Guidelines for Laboratory Testing and Result Reporting of Antibody to Hepatitis C Virus. *MMWR.* 2003;52(Nº RR-3).
24. Busch MP, Lee LL, Satten GA, Henrard DR, Farzadegan H, Nelson KE, et al. Time course of detection of viral and serologic markers preceding human immunodeficiency virus type 1 seroconversion: Implications for screening of blood and tissue donors. *Transfusion.* 1995;35:91.
25. Roca MR, Navas MB, Pignata S. ¿Es conveniente el p24 en donantes? *Rev Arg Transf.* 2001;XXVII Supl:22.
26. Fontana D, Chiera A, García C, Pomo F, Lampuri J, Borrajo G, et al. Tamizaje del Ag p24 del HIV-1 en donantes de sangre. *Rev Arg Transf.* 2001;XXVII Supl:31.
27. Scott E. The Blood Center as an Alternative Test Site for HIV Antibody Testing. *Transfusion.* 1997;37 Supl:85.
28. Muzzio E, Trione N, Soto I, Corti M. Prevalencia de infección por virus de hepatitis C (HCV) en una unidad de internación de pacientes con SIDA. *Actas del 6º Congreso Argentino de SIDA 2003;* p. 58.
29. Kremer L, Treviño E, Domínguez C, Marouch E, Barbás G, Leiva F, et al. Prevalencia de marcadores serológicos de agentes infecciosos en una población de individuos infectados con HIV en Córdoba, Argentina. *Actas del 6º Congreso Argentino de SIDA 2003;* pag. 75.
30. Land K, Davenport P, Nizzi F, Sutor L, Sayers M. An Evaluation of HTLV I/II Confirmed Positive Donors. *Transfusion.* 2004;44 Supl:92A.
31. FDA, Department of Health and Human Services: Application of Nucleic Acid Testing to Blood Borne Pathogens and Emerging Technologies. *OBRR/CBER/FDA Workshop 12/04/2001.*
32. Fiebig EW, Wright DJ, Rawal BD, Garrett PE, Schumacher RT, Peddada L. Dynamics of HIV Viremia and Antibody Seroconversion in Plasma Donors: Implications for Diagnosis and Staging of Primary HIV Infection. *AIDS.* 2003;17:1871.
33. Aprili G, Gandini G, Piccoli P, Franchini M. Detection of an early HIV-1 infection by HIV RNA testing in an Italian blood donor during the preseroconversion window period. *Transfusion.* 2003;43:848.
34. Dodd RY, Notari EP 4th, Stramer SL. Current prevalence and incidence of infectious disease markers and estimated window-period risk in the American Red Cross blood donor population. *Transfusion.* 2002;42:975.
35. Stramer SL, Glynn SA, Kleinman SH, Strong DM, Caglioti S, Wright DJ, et al. Detection of HIV-1 and HCV Infections among Antibody-Negative Blood Donors by Nucleic Acid-Amplification Testing. *N Engl J Med.* 2004;351:760.
36. Kutner JM, Takatu PM, Polite MB, Takatu E, Rosenblit J, Sztterling L, et al. NAT Detection in a Blood Donor at HIV Antigen and Antibody Window Period, First Case Reported in Brazil. *Transfusion.* 2004;44 Supl:93A.
37. Roth WK, Weber M, Buhr S, Drosten C, Weichert W, Sireis W, et al. Yield of HCV and HIV-1 NAT after screening of 3.6 million blood donations in central Europe. *Transfusion.* 2002;42:862.