

Rapid detection of herpes simplex virus-1 from bronchoalveolar lavage fluids using loop-mediated isothermal amplification (LAMP)



Detección rápida del virus herpes simple-1 en lavado broncoalveolar mediante loop-mediated isothermal amplification (LAMP)

Herpes simplex virus (HSV) is a wide spread pathogen with symptoms during infection ranging from completely asymptomatic to a combination of mild symptoms.¹ In critically ill patients HSV can cause life-threatening pulmonary disease.² Cell culture is the gold standard for the diagnosis of HSV, but 2–3 days are needed to detect the cytopathic effect.¹ New technologies, especially polymerase chain reaction (PCR) allows more rapid, albeit more costly, detection of the virus.³ Loop-mediated isothermal amplification (LAMP)⁴ has shown to be comparable with culture and PCR for the detection of HSV in genital lesions.⁵ The aim of this study was to evaluate the efficacy of LAMP compared to conventional culture to detect HSV-1 in bronchoalveolar lavage fluid (BALF).

A total of 50 BALF samples were included from patients with clinically suspected respiratory infection. The samples were routinely cultured in human fibroblast cells and monitored for up to 3 weeks to determine the presence of HSV by cytopathic effect observation and immunofluorescent detection of the antigen using monoclonal antibodies (MicroTrak[®] HSV1/HSV2 Direct Specimen Identification/Typing Test Kit, TrinityBiotech, Ireland).

A set of six previously described primers against the gene encoding the HSV-1 glycoprotein G (gG) were used,⁶ comprising two outer, two inner and two loop primers. The LAMP reaction was performed with 25 µL of the reaction mixture: 2.5 µL mix of primers (final concentration 0.2 µM for outer primers, 1.6 µM for inner primers and 0.4 µM for loop primers), 15 µL Isothermal Master Mix (Optigene, United Kingdom), 2.5 µL PCR grade water and 5.0 µL DNA/sample. The reaction was conducted in a Versant kPCR System (Siemens) at 65 °C for 40 minutes and the LAMP product detected by real time fluorescence. For each sample three reactions were performed: directly from the sample, after boiling 50 µL of sample for 10 minutes and after automatically extracting the DNA using the DSP Virus/Pathogen Midi kit (Qiagen, Germany) in a QIASymphony automated platform (Qiagen, Germany). The specificity of the LAMP assay was evaluated using extracted DNA from 3 positive samples of HSV-2, varicella zoster virus (VZV), Epstein-Bar virus (EBV), cytomegalovirus (CMV) and herpesvirus 6 (HV6). Molecular detection of HSV-1 was performed by real-time quantitative PCR (rtPCR) (HSV1 Q-PCR Alert AmpliMIX, ELITechGroup NANOGEN[®], Italy) with extracted DNA in case of discrepancies between culture and LAMP.

Culture detected 27 positive and 23 samples were negative, whereas LAMP detected HSV-1 in 23 samples in which DNA was extracted and in 21 samples in which LAMP was performed after boiling or directly with the sample. LAMP products were not detected with the DNAs of HSV-2, VZV, EBV, CMV and HV6. After DNA extraction, LAMP only failed to detect HSV-1 in five cases in which rtPCR was also negative (two cases) or presented late threshold cycles (three cases). In one case with negative culture, all variants of LAMP and rtPCR were positive.

Excluding the two samples with negative results for rtPCR and LAMP, the concordance observed between culture and LAMP was 92.0% (44/48, kappa coefficient of 0.834) for LAMP after extracting the DNA and 88.0% (42/48), kappa coefficient of 0.751) for LAMP after boiling or directly with the sample. The sensitivity of LAMP was 88.0% (22/25) after DNA extraction and 80.0%

Table 1

Sensitivity, specificity, positive predictive value and negative predictive value of LAMP (using DNA obtained by extraction or performed after boiling or directly with the sample) for HSV-1 detection from bronchoalveolar samples.

	LAMP		
	Extract	Boiling	Direct
Sensitivity (%; CI95%)	88.0 (67.7–96.8)	80.0 (58.7–92.4)	80.0 (58.7–92.4)
Specificity (%; CI95%)	95.7 (76.0–99.8)	95.7 (76.0–99.8)	95.7 (76.0–99.8)
PPV (%; CI95%)	95.7 (76.0–99.8)	95.2 (74.1–99.8)	95.2 (74.1–99.8)
NPV (%; CI95%)	88.0 (67.7–96.8)	81.5 (61.3–93.0)	81.5 (61.3–93.0)

LAMP: loop-mediated isothermal amplification; PPV, positive predictive value; NPV, negative predictive value; CI95%: 95% confidence interval.

(20/25) when performed after boiling or using directly the sample (Table 1).

The average time to achieve a positive result was 75.7 minutes after DNA extraction (60 min for extraction plus 15.7 min for LAMP), 30.6 min for LAMP after boiling (10 min for boiling plus 20.6 min for LAMP), and 22.8 min for LAMP performed directly with the sample versus three to five days for virus culture.

LAMP assays are emerging as a good and inexpensive alternative for the diagnosis of HSV-1 infections. Previous studies have evaluated LAMP to detect HSV-1 in vitreous,⁷ cutaneous^{5,8} and cerebrospinal fluid samples,⁹ but this is the first study to evaluate the use of a LAMP in respiratory samples. Our results showed very good concordance between culture and LAMP. Performing LAMP directly from the sample saved time and showed similar results compared to LAMP performed in extracted DNA. Discordances with positive culture and negative rtPCR might be explained by the freeze and thaw cycles (DNA was probably degraded). On the other hand, the case with negative culture and positive LAMP and rtPCR could be explained by the virus not being viable. This is not the first time that the results of LAMP performed with and without DNA extraction have shown similar results.⁸ By avoiding the extraction step, LAMP can be performed in short time and is less expensive. The presence of HSV-1 in BALF is related to patient outcomes,¹⁰ including increased mortality. However, few laboratories include the detection of HSV-1 in respiratory samples from critically ill patients thereby leading to possible misdiagnosis of infection. One of the limitations of this study is the lack of analytical sensitivity, but the main objective is to detect those patients with higher viral loads and worse potential outcome. The results of this study show that the LAMP assay can be used as a rapid screening tool to detect HSV-1 in BALF from these patients.

Fundings

This work was supported by Ajut a la Recerca “Clínic-La Pedrera” 2016 (PEP:HB-16-JF-VG-C) and grant 2014SGR0653 from the Departament de Universitats, Recerca i Societat de la Informació de la Generalitat de Catalunya, by the Ministerio de Economía y Competitividad, Instituto de Salud Carlos III, co-financed by European Regional Development Fund (ERDF) “A Way to Achieve Europe,” the Spanish Network for Research in Infectious Diseases (REIPI RD16/0016/0010)

Bibliografía

- Rechenchoski DZ, Faccin-Galhardi LC, Linhares REC, Nozawa C. Herpesvirus: an underestimated virus. *Folia Microbiol (Praha)*. 2017;62:151–6, <http://dx.doi.org/10.1007/s12223-016-0482-7>
- Bruynseels P, Jorens PG, Demey HE, Goossens H, Pattyn SR, Elseviers MM, et al. Herpes simplex virus in the respiratory tract of critical care patients: a prospective study. *Lancet*. 2003;362:1536–41, [http://dx.doi.org/10.1016/S0140-6736\(03\)14740-X](http://dx.doi.org/10.1016/S0140-6736(03)14740-X)

3. Vila J, Zboromyrska Y, Vergara A, Alejo I, Rubio E, Álvarez-Martínez MJ, et al. Molecular diagnostic methods of respiratory infections. Has the scheme diagnosis changed? *Enferm Infecc Microbiol Clin.* 2016;34:40–6, [http://dx.doi.org/10.1016/S0213-005X\(16\)30218-X](http://dx.doi.org/10.1016/S0213-005X(16)30218-X)
4. Notomi T, Okayama H, Masubuchi H, Yonekawa T, Watanabe K, Amino N, et al. Loop-mediated isothermal amplification of DNA. *Nucleic Acids Res.* 2000;28:E63.
5. Sugiyama H, Yoshikawa T, Ihira M, Enomoto Y, Kawana T, Asano Y. Comparison of loop-mediated isothermal amplification, real-time PCR, and virus isolation for the detection of herpes simplex virus in genital lesions. *J Med Virol.* 2005;75:583–7, <http://dx.doi.org/10.1002/jmv.20309>
6. Enomoto Y, Yoshikawa T, Ihira M, Akimoto S, Miyake F, Usui C, et al. Rapid diagnosis of herpes simplex virus infection by a loop-mediated isothermal amplification method. *J Clin Microbiol.* 2005;43:951–5, <http://dx.doi.org/10.1128/JCM.43.2.951-955.2005>
7. Reddy AK, Balne PK, Reddy RK, Mathai A, Kaur I. Loop-mediated isothermal amplification assay for the diagnosis of retinitis caused by herpes simplex virus-1. *Clin Microbiol Infect.* 2011;17:210–3, <http://dx.doi.org/10.1111/j.1469-0691.2010.03216.x>
8. Kobayashi T, Yagami A, Suzuki K, Ihira M, Yoshikawa T, Matsunaga K. Clinical utility of loop-mediated isothermal amplification assay for the diagnosis of common alpha herpesvirus skin infections. *J Dermatol.* 2013;40:1033–7, <http://dx.doi.org/10.1111/1346-8138.12325>
9. Kimura H, Ihira M, Enomoto Y, Kawada J, Ito Y, Morishima T, et al. Rapid detection of herpes simplex virus DNA in cerebrospinal fluid: comparison between loop-mediated isothermal amplification and real-time PCR. *Med Microbiol Immunol.* 2005;194:181–5, <http://dx.doi.org/10.1007/s00430-005-0242-9>
10. Linssen CFM, Jacobs JA, Stelma FF, van Mook WNKA, Terporten P, Vink C, et al. Herpes simplex virus load in bronchoalveolar lavage fluid is related to poor outcome in critically ill patients. *Intensive Care Med.* 2008;34:2202–9, <http://dx.doi.org/10.1007/s00134-008-1231-4>

Andrea Vergara^{a,b,*}, Roser Vendrell^a,
Izaskun Alejo-Cancho^a, Cristina Rodríguez^a, Jordi Vila^{a,b},
M^a Angeles Marcos^{a,b}

^a *Department of Clinical Microbiology, CDB, Hospital Clínic of Barcelona, Barcelona, Spain*

^b *ISGlobal, Barcelona Centre for International Health Research (CRESIB) – Universitat de Barcelona, Barcelona, Spain*

* Corresponding author.

E-mail address: VERGARA@clinic.cat (A. Vergara).

<https://doi.org/10.1016/j.eimc.2018.10.008>
0213-005X/

© 2018 Elsevier España, S.L.U. and Sociedad Española de Enfermedades Infecciosas y Microbiología Clínica. All rights reserved.